

Manual Supplement

Manual Title: 1623-2 Users
Print Date: January 2014
Revision/Date:

Supplement Issue: **2**
Issue Date: 5/15
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 1623-2/1625-2
CD Rev. & Date: 1/2014
CD PN: 4325414

Change #1, 494, 495

On page 30, replace the table with:

Resolution	Range	Accuracy	Operating Error
0.001	0.020 to 1.000 Ω	$\pm (2 \% \text{ rdg} + 10 \text{ digits})$	$\pm (5 \% \text{ rdg} + 3 \text{ digits})$
0.001 to 10 Ω	1.001 to 19.99 k Ω	$\pm (2 \% \text{ rdg} + 3 \text{ digits})$	$\pm (5 \% \text{ rdg} + 3 \text{ digits})$

On page 31, replace both tables with:

Resolution	Range	Accuracy	Operating Error
0.001	0.020 to 1.000 Ω	$\pm (7 \% \text{ rdg} + 10 \text{ digits})$	$\pm (10 \% \text{ rdg} + 5 \text{ digits})$
0.001 to 10 Ω	1.001 to 19.99 k Ω	$\pm (7 \% \text{ rdg} + 3 \text{ digits})$	$\pm (10 \% \text{ rdg} + 5 \text{ digits})$

Resolution	Range	Accuracy	Operating Error
0.001 to 0.1 Ω	0.020 Ω to 90.9 Ω	$\pm (7 \% \text{ rdg} + 3 \text{ digits})$	$\pm (10 \% \text{ rdg} + 5 \text{ digits})$
1 Ω	91 to 199.9 Ω	$\pm (12 \% \text{ rdg} + 3 \text{ digits})$	$\pm (15 \% \text{ rdg} + 5 \text{ digits})$

Change #2, 151

On page 19, replace the 2nd Note with:

Note

Minimum distance between the two current clamps is 15 cm.

On page 20, in Figure 8 change:

From: $\geq 10 \text{ cm}$

To: $\geq 15 \text{ cm}$