

Manual Supplement

Manual Title: 6102 Users Guide
Print Date: January 2013
Part Number: 3729224
Revision/Date:

Supplement Issue: **3**
Issue Date: 6/25
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.

FLUKE®

Calibration

Change #1, 545

On page 11, under **4.1 Unpacking**, remove the 10th bullet.

Change #2, 338

On page 21, under **8.1 Switching to 230V Operation**, remove the 2nd sentence:

Switching the voltage can change the calibration, so *the unit should be calibrated after changing the input voltage.*

Change #3, TPUBS-225

On page 52, under **12.2.1 Compute DELTA**, change these equations

From:

$$\text{delta} = \frac{AF = BE}{DE = CF}$$

To:

$$\text{delta} = \frac{AF - BE}{DE - CF}$$

Under **12.2.2 Compute R0 & ALPHA**, change these equations

From:

$$\text{rzero} = \frac{R_3 a_1 = R_1 a_3}{a_1 - a_3}$$

$$\text{alpha} = \frac{R_1 - R_3}{R_3 a_1 = R_1 a_3}$$

To:

$$\text{rzero} = \frac{R_3 a_1 - R_1 a_3}{a_1 - a_3}$$

$$\text{alpha} = \frac{R_1 - R_3}{R_3 a_1 - R_1 a_3}$$