

Supplement

Title:	700PTP-1 I.S.	Supplement Issue:	4
Part Number:	2811824	Issue Date:	2/25
Print Date:	April 2007	Page Count:	1
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above Instruction Sheet.



PN 6072704

© 2009-2025 Fluke Corporation.
All rights reserved.

FLUKE®

Change #1

Replace the **Replacement Parts** section with:

Hose Assembly, Fluke PN 2815723

Rebuild Kit, Fluke PN 3093412

Bleed Valve Assembly (Needle), Fluke PN 2844329

Handle Assembly Rebuild Kit, Fluke PN 3668409

Change #2, 57806

Under **Introduction**,

Change: Vacuum to -0.96 bar (-13.9 psi)

To: Vacuum to -0.87 bar (-12.7 psi)

Change #3, TPUBS-272

Replace the **Box Contents** section with:

- 700PTP-1 Pneumatic Test Pump
- 300 psi/600 psi adapters:
 - ¼ in male NPT to ¼ in female BSP
 - ¼ in female BSP to ⅜ in female NPT
- 300 psi hose kit:
 - 1 m hose (white)
 - (2) ⅜ in male NPT quick connects (300 psi)
- 600 psi hose kit:
 - 1 m hose (black) with M16X2.0 quick connect fittings
 - (2) ⅜ in male NPT to M16X2.0 quick connect fittings
 - Spare O-rings for 600 psi hose
- Seal Kit
- Instruction Sheet