

Manual Supplement

Manual Title:	805/805 FC Users	Supplement Issue:	2
Print Date:	May 2012	Issue Date:	4/19
Revision/Date:	2, 12/14	Page Count:	1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 638

On page 4, under **Infrared Thermometer**, replace the entire Accuracy (typical) specification with:

Accuracy (typical)..... ± 5 °C from -20 °C to 50 °C, ± 10 °C from 50 °C to 150 °C

Change #2, 646

On page 5, add the following to the **Specifications** section:

Frequency2.4 GHz

Modulation.....GFSK

Output power.....<10 mw