

Manual Supplement

Manual Title: 9142, 9143, 9144 Users Guide
Print Date: February 2013
Part Number: 3720986
Revision/Date:

Supplement Issue: **3**
Issue Date: 11/18
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.



Change #1, 488

On page 13, in the **Specification** table, replace the **TC Millivolt Range** with:

TC Millivolt Range	-10 mV to 100 mV
--------------------	------------------

Change #2, 539

On page 12, in the **Base Unit Specifications** table, replace the **Computer Interface** with:

Computer Interface	RS-232
--------------------	--------

Change #3, 639

On page 5, under **General**, replace the 6th paragraph with:

Before initial use, after transport, or after storage in humid or semi-humid environments, energize the instrument for a "dry-out" period of 2 hours before you assume that the instrument can meet all of the safety requirements of the IEC 61010-2-010. Use a 2 °C/min scan rate and a set point of 100 °C.