

Manual Supplement

Manual Title: 1586A Users
Print Date: June 2013
Revision/Date: Rev. 1, 5/24

Supplement Issue: **1**
Issue Date: 1/26
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.

FLUKE®

Calibration

Change #1, 524

On page 2-3, under **Set the Regional Voltage**, delete:

Each voltage selection requires a specific fuse.

Prior to Table 2-1, add:

Note

See the rear panel decal on the back of your Product to determine the correct fuse.

Replace Table 2-1 with:

Voltage Selector	Fuse	Fluke Part Number
100 V	0.25 A, 250 V (slow blow)	PN 166306
120 V	0.25 A, 250 V (slow blow)	PN 166306
220 V	0.25 A, 250 V (slow blow)	PN 166306
240 V	0.25 A, 250 V (slow blow)	PN 166306