

# Supplement

|                |                         |                   |       |
|----------------|-------------------------|-------------------|-------|
| Manual Title:  | i310s Instruction Sheet |                   |       |
| Part Number:   | 2842344                 | Supplement Issue: | 4     |
| Print Date:    | June 2007               | Issue Date:       | 10/17 |
| Revision/Date: |                         | Page Count:       | 2     |

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.

## Change #1

Under **Battery Replacement**, replace the Note with the following:

*Note*

*For specified battery life, use an alkaline battery.*

## Change #2, 51552

Under the **Specifications**, in the Accuracy section, delete: @25 °C.

## Change #3, 57104

Under **Specifications, Electrical Characteristics**,

Change: All accuracies stated at  $23\text{ °C} \pm 1\text{ °C}$   
( $73.4\text{ °F} \pm 33.8\text{ °F}$ )

To: All accuracies stated at  $23\text{ °C} \pm 1\text{ °C}$

## Change #4, 618

On the first page, add these Symbols to the table:

|                                                                                     |                                                                                                                                                       |
|-------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------|
|  | Consult user documentation.                                                                                                                           |
| <b>CAT III</b>                                                                      | MEASUREMENT CAT III is applicable to test and measuring circuits connected to the distribution part of the building's low-voltage mains installation. |
|  | Conforms to Australian standards.                                                                                                                     |
|  | Battery                                                                                                                                               |

On the 3<sup>rd</sup> page, under **Specifications**, replace the Power Supply with:

|              |                         |
|--------------|-------------------------|
| Power Supply | 9 V Alkaline, IEC 6LR61 |
|--------------|-------------------------|

On the 4<sup>th</sup> page, remove the **Safety Standards** and the **EMC Standards**, sections and replace with:

Safety General.....IEC 61010-1:

Pollution Degree 2

Measurement..... IEC 61010-2-032:

CAT III 300 V

IEC 61326-1: Portable Electromagnetic Environment;  
IEC 61326-2-2

CISPR 11: Group 1, Class A

*Group 1: Equipment has intentionally generated and/or use conductively coupled radio-frequency energy which is necessary for the internal functioning of the equipment itself.*

*Class A: Equipment is suitable for use in all establishments other than domestic and those directly connected to a low voltage power supply network which supplies buildings used for domestic purposes. There may be potential difficulties in ensuring electromagnetic compatibility in other environments, due to conducted and radiated disturbances.*

*Caution: This equipment is not intended for use in residential environments and may not provide adequate protection to radio reception in such environments.*

On the 7<sup>th</sup> page, under **Maintenance, Cleaning**, replace the first sentence with:

Clean the case periodically by wiping it with a damp cloth and mild detergent.

On the 7<sup>th</sup> page, replace the last sentence in the *Note* with:

*Fit only the type 9 V IEC 6LR61*